



**TENDER NOTICE**

**SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT**

**AMREF HEALTH AFRICA**

**TENDER NO./AMREF/02/05/2025/014**

**LOT TWO (2)**

**SUPPLY AND DELIVERY OF BIOCHEMISTRY & ELECTROLYTES ANALYZERS**

## Table of Contents

<b>PART A: TECHNICAL REQUIREMENTS</b>	<b>3</b>
SECTION 1: Invitation to Tender	3
SECTION 2: Instructions to Tenderers	5
2.1 Eligible Tenderers	5
2.2 Cost of tendering	5
2.3 Specific Instructions	5
2.4 Amendment of documents	5
2.5 Tender Prices and Currencies	6
2.6 Validity of Tenders	6
2.7 Sealing and Marking of Tenders	6
2.8 Deadline for Submission of Tenders	6
2.9 Modification of Tenders	6
2.10 Withdrawals of Tender	6
2.11 Opening of Tenders	6
2.12 Clarification of Tenders	7
2.13 Evaluation and Comparison of Tenders	7
2.14 Notification of Award	7
2.15 Appeal Period	7
2.16 Contacting Amref Health Africa	7
2.17 Leadtime and Delivery Details	8
SECTION 3: Eligibility Requirements and Technical Specifications	9
SECTION 4: Confidential Business Questionnaire	15
SECTION 5: Tender Security Form	16
SECTION 6: Manufacturer's Authorization Form	17
SECTION 7: References	18
SECTION 8 Delivery schedule of goods	19
<b>PART B: FINANCIAL REQUIREMENTS</b>	<b>20</b>
SECTION 1: Price Schedule	20
SECTION 2: Delivery Leadtime	20
DECLARATION	22
DELIVERY COMMITMENT FORM	23
ANNEX 1: STAGES OF TENDER	24

## PART A: TECHNICAL REQUIREMENTS

### SECTION 1: Invitation to Tender

1. Amref Health Africa invites sealed bids from eligible candidates for supply and delivery of various laboratory equipment as per the lots listed below;

Lot	Description	UOM	Qty
1	Supply and delivery of Biosafety Cabinets Class II, Type 2A	PC	5
2	Supply and Delivery of Biochemistry & Electrolytes Analyzers	LOT	1
3	Supply and Delivery of Automated Haematology Analyzer	PC	11
4	Supply and Delivery of Integrated Emergency Shower with Eye Wash Station and Laboratory Ergonomic Chairs	LOT	1
5	Supply and Delivery of Medical Laboratory Refrigerators and Freezers	LOT	1
6	Supply and Delivery of Binocular Microscope, Water Distiller and Autoclaves	LOT	1
7	Supply and Delivery of Air Conditioning (ACs) Units	PC	37

2. This tender document is only for **Lot Two (2)**.
3. Interested bidders **MUST** pay a non-refundable fee of **KES 2,000 PER LOT**. The payment should be made through the following Bank account(s): **Kenya Commercial Bank, Kipande House Branch, Account No: 1111429243 OR National Bank of Kenya, Wilson Branch, Account No: 01020058235400** or via **MPesa Paybill 890750, Account 'Lab Equipment'**

**NOTE:** Amref Health Africa will not issue eTIMS receipt for payment made.

4. The bidders will then attach the Mobile money deposit receipt/banking deposit slip as proof of purchase of tender document.
5. Completed tender documents for preliminary and technical requirements are to be enclosed in a plain envelope (**each LOT in a separate envelope**) marked with the tender reference number and tender name. Within the single outer envelope, **the financial bid MUST be in a separate envelope marked with the tender reference number, tender name, vendors name and vendors contact details. The financial bid will only be opened for those bidders who will have qualified in the technical evaluation. Please note that only one complete tender document is required (Do not provide additional copies).**

6. Tenders must be delivered to the address below not later than **Friday 16<sup>th</sup> May 2025 at 12.00 noon** and must be accompanied by a tender security of **KES 648,000.00** in the Currency specified in the tender document in the form of a bank guarantee from a reputable bank and must be delivered with the Tender Documents in the technical bid, **(bid bonds from insurance companies will NOT be accepted)**. The tender security validity period from date of closing tender should also be indicated.
7. Interested eligible bidders are also invited for a virtual pre-tender conference to be held on **Tuesday 06<sup>th</sup> May 2025 starting 10:00am**. Register in advance for this meeting;

<https://amref.zoom.us/meeting/register/qc8w3Ge3S5yPZvJyljfkWw>

After registering, you will receive a confirmation email containing information about joining the meeting.

8. Completed Tender Documents are to be enclosed in a plain sealed envelope(s) clearly marked with the tender name and tender reference number and should be addressed to the following address:

**Group Financial Resources Director  
Amref Health Africa  
P.o. Box 30125-00100,  
NAIROBI**

9. Tender Documents should be received on or before **Friday 16<sup>th</sup> May 2025 at 12.00 noon** at the **Amref Health Africa Kenya Country Office Tender Box** at the **Main Reception**. Electronic bidding will not be permitted. Late tenders will be rejected. Tenders will be opened immediately thereafter in the presence of the 'Tenderers' representatives who choose to attend the tender opening ceremony at the Amref Health Africa Offices.
10. Prices quoted should be inclusive of VAT and all other applicable taxes and must be in Kenya Shillings and shall remain valid for 90 days from the closing date of the tender. The prices in your financial quotation should be broken down as follows: (Do not key in your financial quotation in this section)

- (i) Unit price
- (ii) 16% VAT
- (iii) Other applicable taxes
- (iv) Transport Cost if applicable
- (v) TOTAL COST

**NOTE: Global Fund grants are tax exempt. Bidders will be required to submit quotations inclusive of 16% VAT. Payment(s) to the successful bidder(s) shall be made in total exclusive of VAT and tax exemption certificate will be issued accordingly. Successful applicants shall be required to submit proforma invoice and valid tax compliance certificate upon receipt of PO for processing of specific VAT exemption certificate.**

## **SECTION 2: Instructions to Tenderers**

### **2.1 Eligible Tenderers**

- 2.1.1 This Invitation for Tender is open to all eligible bidders.
- 2.1.2 Tenderers shall not be under a declaration of ineligibility for corrupt or fraudulent practices.

### **2.2 Cost of tendering**

- 2.2.1 The tenderer shall bear all costs associated with the preparation and submission of its bid. Amref Health Africa or its agents, will under no circumstance be responsible or liable for those costs regardless of the conduct or outcome of the tendering process.

### **2.3 Specific Instructions**

- 2.3.1 Bidders must quote for all items and quantities as indicated in the tender document in order to qualify for evaluation.
- 2.3.2 The final bound tender document must be serially paginated (All pages in the document from top page (immediately after top cover) to the last page (one before back cover) including table of content, separators, brochures, bank deposit slip and any other attachments) in a continuous ascending order from the first page to the last in this format; 1,2,3.....n where n is the last numerical page number.
- 2.3.3 Should the reverse side of any leaf have content, it will also be considered as a page and must also be paginated.
- 2.3.4 Any alterations to the page numbering in the bid document must be clearly countersigned by the bidder. Failure to countersign such modifications shall render the bid non-compliant to the pagination requirement clause 2.3.2.
- 2.3.5 Any secondary attachment to a page (such as bank deposit slips/receipts etc.) should be glued onto that page. Stapling of secondary attachments is not allowed and may render your bid non-responsive.
- 2.3.6 Bids will be evaluated in a lot by lot basis
- 2.3.7 The submitted tender document including brochures and any other attachments shall be written in English.
- 2.3.8 Amref effectively checks the validity of the CE certificates (or equivalent documents such as ISO certificates) with issuing authorities and bidders are fully expected to do the same before submission

### **2.4 Amendment of documents**

- 2.4.1 At any time prior to the deadline for submission of tenders, Amref Health Africa for any reasons, whether at its initiative or in response to a clarification requested by a prospective tenderer, may modify the tender documents by amendments.
- 2.4.2 All prospective candidates that have received the tender documents will be notified of the amendment in writing or by post and will be binding on them.
- 2.4.3 In order to allow prospective tenderers reasonable time in which to take the amendment into account in preparing their tenders, Amref Health Africa at its discretion may extend the deadline for the submission of tenders.

## **2.5 Tender Prices and Currencies**

- 2.5.1 The tenderer shall indicate on the appropriate Price Schedule, the unit prices inclusive of all taxes and the total tender price of the items proposed to be purchased under the contract.
- 2.5.2 Prices quoted by the tenderer shall be fixed during the tender validity period and not subjected to variation on any account. A tender submitted with an adjustable price quotation will be treated as non-responsive and will be rejected.
- 2.5.3 The price quoted shall be in **Kenya Shillings**.

## **2.6 Validity of Tenders**

- 2.6.1 Tenders shall remain valid for 90 days after date of tender opening prescribed by Amref Health Africa, pursuant to paragraph 2.10. Tender valid for a shorter period shall be rejected by Amref Health Africa as non-responsive.
- 2.6.2 In exceptional circumstances, Amref Health Africa may solicit the tenderers consent to an extension of the period of validity. The request and the responses thereto shall be made in writing. The tenderer may refuse the request. A tenderer granting the request will not be required nor permitted to modify its tender.

## **2.7 Sealing and Marking of Tenders**

- 2.7.1 The tenderer shall seal the tender and mark it with the tender reference number and tender name and **“DO NOT OPEN BEFORE 12 noon on Friday 16<sup>th</sup> May 2025**

## **2.8 Deadline for Submission of Tenders**

- 2.8.1 Tenders must be received by Amref Health Africa at the address specified not later than **12 noon, on Friday 16<sup>th</sup> May 2025**

## **2.9 Modification of Tenders**

- 2.9.1 The tenderer may modify or withdraw its tender after the tender's submission provided that written notice of the modification, including substitution of withdrawal of the tenders, is received by Amref Health Africa prior to the deadline prescribed for submission of tenders.
- 2.9.2 The tenderer modification or withdrawal notice shall be prepared, sealed, marked, and dispatched in accordance with the provisions of paragraph 2.10:1. A withdrawal notice may be sent by email but followed by a signed confirmation copy, postmarked no later than the deadline for submission of tender
- 2.9.3 No tenderer may be contacted after the deadline for submission of tenders.

## **2.10 Withdrawals of Tender**

- 2.10.1 No tender may be withdrawn in the interval between the deadline for submission of tenders and the expiration of the period of tender validity specified by the tenderer.
- 2.10.2 A bidder who withdraws its tender after the deadline for submission will forfeit its bid security deposit.

## **2.11 Opening of Tenders**

- 2.11.1 Amref Health Africa will open all tenders in the presence of tenderers' representatives who

choose to attend the tender opening ceremony at **12 noon, on Friday 16<sup>th</sup> May 2025** in the location specified in the tender. The tenderers or representatives who are present shall sign a register evidencing their attendance.

2.11.2 The tenderers' names, tender modifications or withdrawals, and the presence or absence of requisite tender security and such other details as Amref Health Africa, at its discretion may consider appropriate, will be announced at the opening.

2.11.3 Amref Health Africa will prepare a tender opening report.

## **2.12 Clarification of Tenders**

2.12.1 To assist in the examination, evaluation and comparison of tenders Amref Health Africa, at its discretion, may ask the tenderer for a clarification of its tender. The request for clarification and the response shall be in writing, and no change in the prices or Substance of the tender shall be sought, offered, or permitted.

2.12.2 Any effort by the tenderer to influence Amref Health Africa in the tender evaluation, tender comparison or contract award decisions may result in the rejection of the tenderers' tender.

## **2.13 Evaluation and Comparison of Tenders**

2.13.1 Amref Health Africa will examine the tenders to determine whether they are complete, whether any computation errors have been made, whether required securities/tender purchase have been furnished, whether documents have been properly signed and whether the tenders are generally in order. After examination, a tender that will be determined to be substantially non-responsive, will be rejected by Amref Health Africa.

2.13.2 Amref Health Africa will evaluate and compare the tenders, which have been determined to be substantially responsive.

2.13.3 Amref Health Africa will notify bidders who will be required to provide samples in good time with specific sample submission details following **finalization of the preliminary evaluation**.

2.13.4 Amref Health Africa will ensure that the submitted samples are catalogued accordingly and correspond to the assigned bid number(s). The tenderers should ensure that all their samples have been captured correctly.

## **2.14 Notification of Award**

2.14.1 Prior to the expiration of the period of tender validity, Amref Health Africa will notify the successful tenderer in writing that the tender has been accepted.

2.14.2 Simultaneously the other tenderers shall be notified that their tenders have been unsuccessful.

## **2.15 Appeal Period**

2.15.1 Any vendor/service provider who wishes to appeal against the outcome of the tender shall do so in writing within 3 working days of the date of the notification/regret letter. Any letter received after the third day shall not be responded to and shall be treated as null and void. Amref Health Africa shall have dispensed with this procurement.

## **2.16 Contacting Amref Health Africa**

2.16.1 No tenderer shall contact Amref Health Africa on any matter relating to its tender, from the time of the tender opening to the time the contract is awarded.

2.16.2 Any effort by a tenderer to influence Amref Health Africa in its decisions on tender evaluation, tender evaluation committee, or contract award will result in the rejection of the tenderer's tender.

## **2.17 Leadtime and Delivery Details**

2.17.1 The supplier should be able to deliver the items in the lot(s) they have quoted for as specified in this tender document.

2.17.2 The supplier should include delivery schedule with lead times.

2.17.3 On arrival the supplies should be free from damage. The supplier shall be liable for all losses due to insufficient or unsuitable packing and delivery arrangements, and shall be liable for the cost of returning any unacceptable supplies.

2.17.4 The supplies must be free from objectionable matter and any substances that would represent a hazard to health.



### **SECTION 3: Eligibility Requirements and Technical Specifications**

#### **a) Preliminary Evaluation Criteria**

Bids will be evaluated based on the below criteria.

Bids **lacking any of the documents** below will be considered as non-responsive and therefore will be eliminated at this stage.

#### **PRELIMINARY EVALUATION OF MANDATORY REQUIREMENTS**

<b>PRELIMINARY EVALUATION OF MANDATORY REQUIREMENTS</b>				
<b>Mandatory Requirements</b>				
<b>No.</b>	<b>Particulars</b>	<b>Marks</b>	<b>Compliant</b>	<b>Non-compliant</b>
1.	Copy of Certificate of Incorporation/Certificate of Registration	1 or 0		
2.	Copy of valid KRA Tax Compliance certificate	1 or 0		
3.	Must attach Mobile money deposit receipt/ banking deposit slip as proof of purchase of tender documents <b>(If applying for more than one LOT, provide specific payment for each LOT)</b>	1 or 0		
4.	Must provide a copy of valid business Trading License	1 or 0		
5.	Submit a current CR12 from the Registrar of Companies (generated not earlier than <b>March 2024</b> ) or copies of identification cards for Sole Proprietors/Partnerships	1 or 0		
6.	Must provide tender security from a reputable bank amounting to <b>KES 648,000.00</b> (Tender security must be attached per lot) valid until <b>Wednesday 12<sup>th</sup> November 2025</b>	1 or 0		
7.	Must submit copies of the most recent 2 sequential audited financial reports (not older than 2022) and each must be signed and dated by the firms Directors and by the auditor <b>(Both reports must have unqualified/unmodified auditor's opinion).</b>	1 or 0		
8.	<b>ENSURE</b> that all pages are sequentially paginated in the format 1, 2,3 .... starting with 1 on the top page <b>(see details of pagination and binding on sections 2.3.1 to 2.3.5 of the tender document).</b>	1 or 0		

**Note:** All the above documents numbered 1 to 8 should be packaged and arranged in that order under the preliminary evaluation criteria section of the tender document.

a) **Technical evaluation criteria**

Detailed technical Specifications/ Description of requirements	Max	Score
<b>Compliance with Technical Specifications for the following items; (Attach brochures, and/or manuals from the manufacturer to support the specifications below)</b>		
<p>1 Provided Specifications for the items below;</p> <p style="text-align: center;"><b>Biochemistry Analyzer</b></p> <p><b>General Description</b> Fully automated Biochemistry analyzers measure biochemical indices by analyzing blood and other body fluids for disease diagnosis, evaluation of organ functions, identify disease genes and monitor therapeutic efficacy. It should be capable of measuring the following parameters;</p> <ul style="list-style-type: none"> <li>i) Specific proteins</li> <li>ii) Electrolytes</li> <li>iii) Enzymes</li> <li>iv) Substrates</li> <li>v) Drugs of abuse</li> <li>vi) Hormones including TSH (thyroid-stimulating hormone, FSH (Follicle Stimulating Hormone),</li> <li>vii) Cancer biomarkers</li> <li>viii) Therapeutic drugs</li> </ul> <p style="text-align: center;"><b>Performance Specifications</b></p> <p><b>Main Unit</b>  <b>Test menu:</b> At least 120 selections <b>(1 Mark)</b>  <b>Test per hour:</b> 200-400 tests/hour <b>(1 Mark)</b>  <b>On- board parameters:</b> At least 50 <b>(1 Mark)</b>  <b>Analyzer system:</b> Random access, Discrete, Automatic, Selectable STAT Sample Priority <b>(1 Mark)</b>  <b>Programming:</b> User defined and calculations <b>(1 Mark)</b></p> <p><b>Sample processing</b>  <b>Sample tray capacity:</b> 20 to 80 samples <b>(1 Mark)</b>  <b>Sample Volume:</b> Minimum 2µL – Maximum 70 µL <b>(1 Mark)</b>  <b>Sample dilution:</b> Automatic, Pre and post dilutions <b>(1 Mark)</b>  <b>Sample identification:</b> Bar code reader <b>(1 Mark)</b>  <b>Sample probe:</b> Probe crash protection, Liquid level Detection and clot detection <b>(1 Mark)</b>  <b>Probe cleaning:</b> Internal and External <b>(1 Mark)</b>  <b>Sample dead volume:</b> Maximum 100 µL in sample cup and maximum of 50 µL pediatric <b>(1 Mark)</b>  <b>Quality Control (QC) tracking:</b> MUST have automated Levy Jennings charting system for all analytes <b>(1Mark)</b></p> <p><b>Reagent processing</b>  <b>Reagent tray Capacity:</b> 10-60 <b>(1 Mark)</b>  <b>Reagent storage:</b> Refrigerated compartment <b>(1 Mark)</b>  <b>Reagent probe:</b> Probe crash protection, liquid level detection <b>(1 Mark)</b></p>	42 or 0	

<p><b><u>Optical Characteristics</u></b>  <b>Light source:</b> Halogen- Tungsten lamp or the equivalent technology <b>(1 Mark)</b>  <b>Wavelength:</b> 340- 800 nm <b>(1 Mark)</b>  <b>Sensitivity:</b> 0.0001 OD or better <b>(1 Mark)</b></p> <p><b><u>Data Processor</u></b>  <b>Operating system:</b> Compatible with Windows 10 and above <b>(1 Mark)</b>  <b>Interface:</b> RS-232 or USB or Bluetooth <b>(1 Mark)</b>  <b>Memory:</b> &gt; 10GB <b>(1 Mark)</b>  <b>Display:</b> Digital display <b>(1 Mark)</b>  <b>Control and Calibration:</b> Automatic, can be adjusted through the software with controls and calibration equipment proof of traceability <b>(1Mark)</b>  <b>Parameters:</b> Display of running status, alerts <b>(1 Mark)</b></p> <p><b><u>Physical characteristics</u></b>  <b>Main unit:</b> Island bench or bench top <b>(1 Mark)</b></p> <p><b><u>Operating environment</u></b>  <b>Power Requirements:</b> 220-240V, A/c 50-60 Hz <b>(1 Mark)</b>  <b>Ambient temperature of operation:</b> 10° C to 40° C <b>(1 Mark)</b>  <b>Relative humidity:</b> 20% to 90% <b>(1 Mark)</b></p> <p><b><u>Accessories (To be supplied alongside each Biochemistry Analyzer)</u></b>  UPS (Attach brochure) <ul style="list-style-type: none"> <li>• <b>Capacity:</b> 3000VA or more <b>(1 Mark)</b></li> <li>• <b>Input:</b> Ac 100-240V, 50-60Hz <b>(1 Mark)</b></li> <li>• <b>Output:</b> Ac 220-240V, 50-60Hz <b>(1 Mark)</b></li> <li>• <b>Power cable:</b> 3-pin UK plug <b>(1 Mark)</b></li> </ul> <p><b><u>Consumables/Reagents</u></b>  Must provide a commitment letter to provide the manufacturer's recommended set of consumables/ reagents (specify composition) for startup and operation of the system for at least 100 tests with associated consumables <b>(1 Mark)</b></p> <p><b><u>Quality standards</u></b>  <b>Manufacturing Standards:</b> MUST have CE Mark and IEC 60601-1 or UL3101-1, or 13485 or ISO 9001 or any other internationally recognized standards <b>(1 Mark)</b>  <b>Conformity to Standards:</b> CE mark, IVD-Directive 98/79/EC or any other internationally recognized <b>(1 Mark)</b></p> <p><b><u>Installation, testing, and commissioning</u></b>  Provide a detailed workplan for delivery, installation, testing, training, and commissioning of the machines at various sites as per manufacturer's instructions <b>(3 Marks)</b></p> </p>		
--	--	--

	<p><b><u>Manufacturer's Authorization</u></b> Must provide Manufacturer's certificate (if bidding as the manufacturer) or Manufacturer's authorization certificate (if not the manufacturer) <b>(2 Marks)</b></p> <p><b><u>Warranty</u></b> Minimum of one -year replacement warranty in the event of any component failure arising from defective design, material or workmanship after commissioning on the equipment</p> <ul style="list-style-type: none"> <li>• Provided <b>(1 Mark)</b></li> <li>• Less than one year or not provided <b>(0 Marks)</b></li> </ul>		
2	<p>Provided Specifications for the items below;</p> <p style="text-align: center;"><b>Electrolyte Analyzer</b></p> <p><b><u>General Description</u></b> To be used in laboratory for electrolyte estimation/measurement using ion selective electrode method. It should have an external printer system compatible with computer windows</p> <p style="text-align: center;"><b>Performance Specifications</b></p> <p><b><u>Main Unit</u></b>  <b>Measurement range:</b> 0 to 300 mmol/l <b>(1 Mark)</b>  <b>Accuracy:</b> <math>\pm 1\%</math> <b>(1 Mark)</b>  <b>Analysis time:</b> 0-1 Minute <b>(1 Mark)</b>  <b>Display:</b> Digital Display <b>(1 Mark)</b>  <b>Electrodes:</b> Available and separate for each element (Sodium, Potassium, Chloride) <b>(1 Mark)</b>  <b>Sample volume:</b> 0-200<math>\mu</math>L <b>(1 Mark)</b>  <b>Connectivity:</b> USB port <b>(1 Mark)</b>  Wireless or Ethernet <b>(1 Mark)</b>  <b>Printer:</b> In built printer with provision to connect an external printer <b>(1 Mark)</b></p> <p><b><u>Physical Characteristics</u></b>  <b>Main unit:</b> Bench top <b>(1 Mark)</b></p> <p><b><u>Operating Environment</u></b>  <b>Power Requirements:</b> 220-240 V and 50-60Hz <b>(1 Mark)</b>  <b>Ambient temperature:</b> 15° C to 40° C <b>(1 Mark)</b>  <b>Ambient operating humidity:</b> 20% to 90% <b>(1 Mark)</b></p> <p><b><u>Additional Equipment to be Provided</u></b>  UPS  <b>Capacity:</b> 3000VA or more <b>(1 Mark)</b>  <b>Input:</b> Ac100-240V, 50-60Hz, <b>(1 Mark)</b>  <b>Output:</b> Ac 100-240V, 50-60Hz, <b>(1 Mark)</b>  <b>Power Cable:</b> 3-pin UK plug <b>(1 Mark)</b></p>	30 or 0	

	<p><b><u>Consumables/Reagents</u></b></p> <p>Must provide a commitment letter to deliver an electrode module (Start-up kit of 100 tests for each electrolyte) at least for starter alongside the equipment. Controls and Calibrators must also be provided for all these tests <b>(3 Marks)</b></p> <p><b><u>Quality Standards</u></b></p> <p><b>Manufacturing Standards:</b> IEC 60601-1 or 13485 or ISO 9001 or any other internationally recognized Standards <b>(1Mark)</b></p> <p><b>Conformity to Standards:</b> CE mark, IVD-Directive 98/79/EC, or any other internationally recognized documents <b>(1 Mark)</b></p> <p><b><u>Installation, Testing, and Commissioning</u></b></p> <p>Provide a detailed workplan for delivery, installation, testing, training, and commissioning of the machines at various sites as per manufacturer's instructions <b>(3 Marks)</b></p> <p><b><u>Manufacturer's Authorization</u></b></p> <p>Must provide Manufacturer's certificate (if bidding as the manufacturer) or Manufacturer's authorization certificate (if not the manufacturer) <b>(2 Marks)</b></p> <p>Provide a commitment letter to deliver alongside the equipment 2 physical sets of the technical documentations (that is user manual and service manual) in English per site at the time of commissioning <b>(2 Marks)</b></p> <p><b><u>Warranty</u></b></p> <p>Minimum of one -year replacement warranty in the event of any component failure arising from defective design, material or workmanship after commissioning on the equipment</p> <ul style="list-style-type: none"> <li>• Provided <b>(1 Mark)</b></li> <li>• Not provided <b>(0 Marks)</b></li> </ul>		
3	<p>For Biochemistry and Electrolyte Analyzer, <b>provide a commitment letter detailing;</b></p> <ol style="list-style-type: none"> <li>1. Commitment to supply the Biochemistry and Electrolyte Analyzer equipment from the same manufacturer for standardization of reagents and integrated equipment maintenance <b>(3 Marks)</b></li> <li>2. Software support (upgrade) for a minimum of 1 years after installation and any applicable software licensing fees disclosure <b>(3 Marks)</b></li> <li>3. Ability to integrate with other systems (including Laboratory Information System) using electronic communication standards (e.g ASTM, HL7, FHIR) <b>(3 Marks)</b></li> <li>4. Proof of in-house or manufacturer's local presence including proof of adequate facilities locally to offer the service, spare parts, and reagents and qualified and skilled technical staff <b>(3 Marks)</b></li> <li>5. To deliver alongside the equipment 2 physical sets of the technical documentations (that is user manual and service manual) in English per site at the time of commissioning <b>(3 Marks)</b></li> </ol>	15	

4	Indicate lead time for delivery of the equipment from contracting to commissioning; a) 6 weeks or less (10 marks) b) 7-8 weeks (8 marks) c) 9-10 weeks (5 marks) d) More than 10 weeks (1 marks) e) Not Provided (0 Marks)	10	
5	Provide proof of supply and delivery of laboratory equipment in the recent past (2022 to date) specifying the value ( <b>purchase orders/invoices/ contracts</b> ) <ul style="list-style-type: none"> <li>● Orders above Kshs. 20M – 10 marks</li> <li>● Orders above Kshs 10M - but below Kshs 20M- 5 marks</li> <li>● Orders of Kshs 10M or below – 2 marks</li> <li>● Not Provided – 0 marks</li> </ul>	10	
	<b>TOTAL</b>	<b>107 MKS</b>	
	<b>Only bidders who will score at least 97 marks and above shall proceed to the next stage of evaluation</b>		

#### **SECTION 4: Confidential Business Questionnaire**

You are requested to give the particulars indicated in part 1 and either part 2(a), 2(b) or 2(c) whichever applies to your type of business.

##### **Part 1 - General**

Business Name: ..... Location of Business premises:  
Country/Town..... Postal Address: .....  
Code: .....Town: .....  
Tel No.: ..... E-mail: .....  
Fax: ..... Nature of Business: .....  
.....

##### **Part 2 (a) – Individuals**

Your Name in full: ..... Nationality: .....  
Country of Origin: ..... Citizenship details: .....

##### **Part 2 (b) – Partnership**

	Name:	Nationality:	Citizenship Details:	Shares
1.	.....	.....	.....	.....
2.	.....	.....	.....	.....
3.	.....	.....	.....	.....

##### **Part 2 (c) – Registered Company**

Private or Public:  
.....  
State the nominal and issue capital of the company:  
.....  
.....  
Nominal KES:  
.....  
Issued KES:  
.....

Give details of all directors as follows:

	Name:	Nationality:	Citizenship Details:	Shares
1.	.....	.....	.....	.....
2.	.....	.....	.....	.....
3.	.....	.....	.....	.....

## **SECTION 5: Tender Security Form**

**TENDER NO./AMREF/02/05/2025/014 – LOT 2**

### **SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT**

**To:** Amref Health Africa,

WHEREAS *[insert: **name of Tenderer**]* (hereinafter called “the Tenderer”) has submitted its tender dated *[insert: **date of tender**]* for the performance of the above-named Contract (hereinafter called “the Tender”)

KNOW ALL PERSONS by these present that WE *[insert: **name of bank**]* of *[insert: **address of bank**]* (hereinafter called “the Bank”) are bound unto *[insert: **name of Purchaser**]* (hereinafter called “the Purchaser”) in the sum of: *[insert: **amount**]*, for which payment well and truly to be made to the said Purchaser, the Bank binds itself, its successors and assigns by these presents.

Sealed with the Common Seal of the said Bank this *[insert: **number**]* day of *[insert: **month**]*, *[insert: **year**]*.

THE CONDITIONS of this obligation are the following:

1. If, after the tender submission deadline, the Tenderer
  - i. withdraws its tender during the period of tender validity specified by the Tenderer in the Tender Form, or
  - ii. does not accept the Purchaser’s corrections of arithmetic errors in accordance with the Instructions to Tenderers; or
  - iii. does not at all reply to the Purchaser’s requests for clarification.
2. If the Tenderer, having been notified of the acceptance of its tender by the Purchaser during the period of tender validity
  - (a) fails or refuses to sign the Contract Agreement when required; or
  - (b) Fails or refuses to issue the performance security in accordance with the Instructions to Tenderers.

We undertake to pay to the Purchaser up to the above amount upon receipt of its first written demand, without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due it, owing to the occurrence of any one of the two above-named CONDITIONS, and specifying the occurred condition or conditions.

This guarantee will remain in full force up to and including **Wednesday 12<sup>th</sup> November 2025** and any demand in respect thereof must reach the Bank not later than the above date.

**Note: A bidder who withdraws its tender after the deadline for submission will forfeit its bid security deposit.**

For and on behalf of the Bank

Signed: \_\_\_\_\_ Date: \_\_\_\_\_



## SECTION 6: Manufacturer's Authorization Form

*(Must be on Manufacturer's or Producer's letterhead)*

To. Amref Health Africa,  
Opposite Langata Road Primary School,  
P.O Box 30145-00100  
Nairobi, Kenya

Dear Sirs,

WHEREAS *[ insert: name of the manufacturer or producer]* (hereinafter, "we" or "us") who are established and reputable manufacturers or producers of *[ insert: name and/or description of the Goods requiring this authorization]* (hereinafter, "Goods") having production facilities at *[ insert: address of factory]* do hereby authorize *[ insert: name and address of Tenderer]* (hereinafter, the "Tenderer") to submit a tender, and subsequently negotiate and sign the Contract with you against *[ insert: the specific tender name and tender number]* For supply of laboratory equipment, including the above Goods produced by us.

We hereby extend our full guarantee and warranty for the above specified Goods against these tender documents.

For and on behalf of the Manufacturer or Producer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

In the capacity of *[insert: title, position, or other appropriate designation]* and duly authorize to sign this Authorization on behalf of *[insert: name of manufacturer or producer]*

NOTE: Manufacturer's Authorization must be ON LETTER HEAD and addressed to Amref Health Africa and must be specific to the tender and item and signed by an authorized signatory.  
- MANDATORY

## **SECTION 7: References**

<b>BANK REFERENCES</b>	
<b>BANK HOLDING MAIN ACCOUNT</b>	
Bank name and address:	
Name of account:	
Account number:	How long open?

<b>COMMERCIAL REFERENCES</b>	
Provide names and contract details of two customers who may be approached to verify your capacity to perform against similar contracts.	
<b>LOCAL/INTERNATIONAL TRADE REFERENCE – CUSTOMER 1</b>	
Name and address:	
Activity:	Period of relationship:
Contact name:	Fax no.:
Telephone No.:	
<b>LOCAL/INTERNATIONAL TRADE REFERENCE – CUSTOMER 2</b>	
Name and address	
Activity:	Period of relationship:
Contact name:	Fax no.:
Telephone No.:	

## SECTION 8 Delivery schedule of goods

DESCRIPTION	QTY	UNIT OF MEASURE	DELIVERY LOCATION
Biochemistry Analyzer	12	PC	St. Theresa Kiirua Nanyuki CRH Narok CRH Kwale SCH Samburu CRH Taita Taveta SCH Makueni CRH Moyale CRH Isiolo CRH Garissa CRH Karatina SCH National Laboratory Services
Electrolyte Analyzer	12	PC	St. Theresa Kiirua Nanyuki CRH Narok CRH Kwale SCH Samburu CRH Taita Taveta SCH Makueni CRH Moyale CRH Isiolo CRH Garissa CRH Karatina SCH National Laboratory Services

**Note: Bidders are to quote for supply, delivery, and installation of the equipment to each facility as listed above**

## **PART B: FINANCIAL REQUIREMENTS**

**PLEASE PROVIDE THIS IN A DIFFERENT ENVELOPE** [The financial bid **MUST** be in a separate envelope marked with the tender reference number, tender name, vendors name and vendors contact details. The financial bid will only be opened for those bidders who will have qualified in the technical evaluation]

### **SECTION 1: Price Schedule**

**(Quote for each item as detailed in the delivery schedule of goods)**

Lot No	TENDERERS NAME	SPECIFICATIONS	TENDER NUMBER	QTY	UNIT PRICE	VAT	TOTAL	REMARKS
2								
		<b>TOTAL</b>						

### **SECTION 2: Delivery Leadtime**

Lot No	TENDERERS NAME	SPECIFICATIONS	TENDER NUMBER	QTY	DELIVERY LEAD TIME	REMARKS
2						
		<b>TOTAL</b>				

**Note. In case of discrepancy between the unit price and total, the unit price shall prevail.**

<b>Currency:</b>	<b>Grand Total Bid Price:</b>	<b>Amount in Figures:</b>
		<b>Amount in Words</b>
<b>Bidder's Name and Address</b>	<b>Date</b>	<b>Signature and Stamp</b>

**Note: Indicate breakdown of all taxes applicable.**

**DECLARATION**

I/We have completed this form (s) accurately at the time of reply and it is agreed that all responses can be substantiated, if requested to do so, and that any inaccuracy in the information filled herein will lead to disqualification of the tenderer.

For and behalf of:

.....

Name:

.....

Date: ..... Signature:

.....

**DELIVERY COMMITMENT FORM**

I/We *[insert: tenderers name]* acknowledge the delivery schedule above for the procurement of *[insert: description]* and do hereby commit ourselves that we shall deliver the goods within *[insert: timelines]* as stipulated in this tender document.

For and behalf of:

..... Name:  
.....

Date: ..... Signature: .....

## **ANNEX 1: STAGES OF TENDER**

### **PART A**

#### **a) Preliminary Evaluation**

Tenderers are required to comply with mandatory requirements

- ✓ Bidders who shall not provide any of the documents shall be considered non-responsive and shall not proceed to the next stage(s) of evaluation

#### **b) Technical Evaluation**

##### **i. Documents Examination**

- ✓ Bidders who will not meet the pass mark(s) as specified shall be considered non-responsive and shall not proceed to the next stage(s) of evaluation

##### **ii. Product Evaluation**

- ✓ The technical evaluation shall involve the product evaluation and or samples where applicable

### **PART B**

#### **a) Financial Evaluation & Delivery lead time evaluation**

- ✓ Tenderers who are successful at proceeding stages shall have their prices and delivery period compared and award recommended to the lowest evaluated responsive bid.

#### **b) Contracting**

- ✓ If accepting of the offer, the successful bidder shall be contracted per the sample standard agreement accessible on the Amref website.